

# United States Patent and Trademark Office

UNITED STATES DEPARTMENT OF COMMERCE United States Parent and Trademark Office address: Commissioner For Patents P. d. Bo. 1450 Advandria, Virginia 22313-1450

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/622,720	07/18/2003	Richard Rollin	1801/91	2950
7590 04/21/2006			EXAMINER	
MICHAEL H. BANIAK			ANDERSON, CATHARINE L	
BANIAK PINE	& GANNON		A DT LINET	PAPER NUMBER
Suite 1200			ART UNIT	FAPER NUMBER
150 N. Wacker Drive Chicago, IL 60606			3761	
			DATE MAILED: 04/21/2006	

Please find below and/or attached an Office communication concerning this application or proceeding.

	Application No.	Applicant(s)			
		ROLLIN, RICHARD			
Office Action Summary	10/622,720				
• • • • • • • • • • • • • • • • • • •	Examiner	Art Unit			
The MAILING DATE of this communication and	C. Lynne Anderson	3761			
The MAILING DATE of this communication appears on the cover sheet with the correspondence address Period for Reply					
A SHORTENED STATUTORY PERIOD FOR REPLY WHICHEVER IS LONGER, FROM THE MAILING DATE - Extensions of time may be available under the provisions of 37 CFR 1.13 after SIX (6) MONTHS from the mailing date of this communication.  If NO period for reply is specified above, the maximum statutory period was realized to reply within the set or extended period for reply will, by statute, Any reply received by the Office later than three months after the mailing earned patent term adjustment. See 37 CFR 1.704(b).	ATE OF THIS COMMUNICATION 36(a). In no event, however, may a reply be tin vill apply and will expire SIX (6) MONTHS from 1, cause the application to become ABANDONE	N. nely filed the mailing date of this communication. D (35 U.S.C. § 133).			
Status					
1) Responsive to communication(s) filed on 18 July 2003.					
, <del>_</del>	· · · · · · · · · · · · · · · · · · ·				
3) Since this application is in condition for allowance except for formal matters, prosecution as to the merits is					
closed in accordance with the practice under Ex parte Quayle, 1935 C.D. 11, 453 O.G. 213.					
Disposition of Claims					
4) ☐ Claim(s) 1 is/are pending in the application.  4a) Of the above claim(s) is/are withdray  5) ☐ Claim(s) is/are allowed.  6) ☐ Claim(s) 1 is/are rejected.  7) ☐ Claim(s) is/are objected to.  8) ☐ Claim(s) are subject to restriction and/or		•			
Application Papers					
9) The specification is objected to by the Examine 10) The drawing(s) filed on is/are: a) access applicant may not request that any objection to the Replacement drawing sheet(s) including the correct 11) The oath or declaration is objected to by the Examine	epted or b) objected to by the drawing(s) be held in abeyance. Se ion is required if the drawing(s) is ob	e 37 CFR 1.85(a). jected to. See 37 CFR 1.121(d).			
Priority under 35 U.S.C. § 119					
<ul> <li>12) Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).</li> <li>a) All b) Some * c) None of:</li> <li>1. Certified copies of the priority documents have been received.</li> <li>2. Certified copies of the priority documents have been received in Application No.</li> <li>3. Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).</li> <li>* See the attached detailed Office action for a list of the certified copies not received.</li> </ul>					
Attachment(s)  1) Notice of References Cited (PTO-892)  2) Notice of Draftsperson's Patent Drawing Review (PTO-948)  3) Information Disclosure Statement(s) (PTO-1449 or PTO/SB/08) Paper No(s)/Mail Date 3/29/04.	4) Interview Summary Paper No(s)/Mail D 5) Notice of Informal I 6) Other:				

Application/Control Number: 10/622,720

Art Unit: 3761

## **DETAILED ACTION**

# Claim Rejections - 35 USC § 102

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless -

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

Claim 1 is rejected under 35 U.S.C. 102(b) as being anticipated by Silver (5,797,875).

With respect to claims 1, 6, 35, and 40, Silver discloses an adapter 32, as shown in figure 1, connecting a vacuum source 24 to a breast shield 22. The adapter 32 comprises a housing 46, as shown in figure 2, having two female coupling components 52 defined by the interior sidewall surface of the housing. A port 42 allows communication with the internal chamber of the housing via channels 59. The female coupling components 52 are sized and shaped to fit snugly with the male coupling components, as disclosed in column 3, lines 62-66. The male coupling components 54, as shown in figure 22, have a first end, a second end, and a passageway connecting the ends. The first end receives vacuum tubing 44 and the second end receives the female coupling component, as disclosed in column 5, lines 52-61. The second end has a first end diameter 82, a second end diameter 86 that is greater than the first end diameter 82, and a sealing surface 84. The male coupling component is received in the upper portion of the female coupling component to form an airtight engagement. Since the female coupling components 52 are sized and shaped to fit snugly with the male

Art Unit: 3761

coupling components 54, the female components therefore have an upper portion having a diameter matching the second end of the male component, and a lower portion having a diameter matching the first end of the male component. Therefore, the female coupling components have an upper portion having a diameter greater than the diameter of the lower portion. A stopper 70 is provided to engage the upper portion of the female coupling component, as shown in figure 21. An aperture 63 extends between ambient atmosphere and the female coupling component, as disclosed in column 5, lines 30-33.

### **Double Patenting**

The nonstatutory double patenting rejection is based on a judicially created doctrine grounded in public policy (a policy reflected in the statute) so as to prevent the unjustified or improper timewise extension of the "right to exclude" granted by a patent and to prevent possible harassment by multiple assignees. A nonstatutory obviousness-type double patenting rejection is appropriate where the conflicting claims are not identical, but at least one examined application claim is not patentably distinct from the reference claim(s) because the examined application claim is either anticipated by, or would have been obvious over, the reference claim(s). See, e.g., *In re Berg*, 140 F.3d 1428, 46 USPQ2d 1226 (Fed. Cir. 1998); *In re Goodman*, 11 F.3d 1046, 29 USPQ2d 2010 (Fed. Cir. 1993); *In re Longi*, 759 F.2d 887, 225 USPQ 645 (Fed. Cir. 1985); *In re Van Ornum*, 686 F.2d 937, 214 USPQ 761 (CCPA 1982); *In re Vogel*, 422 F.2d 438, 164 USPQ 619 (CCPA 1970); and *In re Thorington*, 418 F.2d 528, 163 USPQ 644 (CCPA 1969).

A timely filed terminal disclaimer in compliance with 37 CFR 1.321(c) or 1.321(d) may be used to overcome an actual or provisional rejection based on a nonstatutory double patenting ground provided the conflicting application or patent either is shown to be commonly owned with this application, or claims an invention made as a result of activities undertaken within the scope of a joint research agreement.

Effective January 1, 1994, a registered attorney or agent of record may sign a terminal disclaimer. A terminal disclaimer signed by the assignee must fully comply with 37 CFR 3.73(b).

Claim 1 is provisionally rejected on the ground of nonstatutory obviousness-type double patenting as being unpatentable over claims 7 and 35 of copending Application No. 10/600,078. Although the conflicting claims are not identical, they are not

Application/Control Number: 10/622,720 Page 4

Art Unit: 3761

patentably distinct from each other because it would be obvious that the amount of flow of the stoppered coupling component is less than the flow of the unstoppered coupling component since the purpose of the stopper is to impede flow.

This is a <u>provisional</u> obviousness-type double patenting rejection because the conflicting claims have not in fact been patented.

#### Conclusion

The prior art made of record and not relied upon is considered pertinent to applicant's disclosure. U.S. Patent 6,706,012 and U.S. Patent Application Publication US 2001/0038799 A1 disclose adapters for use with vacuum sources in breast sheilds.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to C. Lynne Anderson whose telephone number is (571) 272-4932. The examiner can normally be reached on Monday through Friday.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Tanya Zalukaeva can be reached on (571) 272-1115. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Application/Control Number: 10/622,720 Page 5

Art Unit: 3761

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

(W4cla April 17, 2006

TATYANA ZALUKAEVA BUPERVISORY PRIMARY EXAMINER